



Date:

Report Type:

2023/5/22

Maintenance

Customer:

Name: Dr. Sagar Sami	Location: Sul
----------------------	---------------

Machine & Inspection:

Name	Type	Model	Serial Number	Lamp Shoots	✓/✗
switch	nd:gag	gag 2c	2022 01240702	88,183	✓

Report:

The HP. connector was danged.

Actions:

I Replace it. (both)

Price:

\$ 200

Engineer name & signature:

[Signature]

Customer Name & signature

[Signature]